



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor P. Zippo
Location 28 Pct.
Date of Inspection 6/18/08

Intoxilyzer Serial # 68-011473 Simulator Serial # DR 3556 Lot # 08010
1. .096
2. .097
3. .097
Diagnostic test OK
Breath hose heated OK
Printer check OK

Intoxilyzer Serial # 68-010779 Simulator Serial # DR 4153 Lot # 08010
1. .104
2. .105
3. .105
Diagnostic test OK
Breath hose heated OK
Printer check OK

REMARKS:

Intox. Ser. # 68-010779 found 01/05. Performed field inspection, working properly. Ret'd back to Service - Hwy. Dist. not field.

028 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
06/18/2008

DIAGNOSTIC TEST 16:29 EDT

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240

RAM CHECK PASSED
TEMP CHECK PASSED

PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMNOPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

028 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
06/18/2008

TEST	%BrAC	TIME
AIR BLANK	.000	16:30EDT
CAL. CHECK	.104	16:30EDT
AIR BLANK	.000	16:30EDT
CAL. CHECK	.105	16:31EDT
AIR BLANK	.000	16:31EDT
CAL. CHECK	.105	16:31EDT
AIR BLANK	.000	16:31EDT
NO. OF SAMPLES:	03	
MEAN:	.1046	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679 - 1990

SN 68-010779
SLAVE 75 2240
DATE 06/18/2008

VERSION G1776.12

TIME 16:32 EDT

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^*()

028 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679 - 1990

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990



"B"

6/18/08
68-010779
Found o/p/s,
instrument
powered off.

(100)

6/14/08 8:00 am