




9
POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor P. Zippo
Location 028 Pct. 
Date of Inspection 6/19/08

Intoxilyzer Serial # 68013546 Simulator Serial # DR4147 Lot # 08010
1. .101
2. .103
3. .102
Diagnostic test OK
Breath hose heated OK
Printer check OK

Intoxilyzer Serial # _____ Simulator Serial # _____ Lot # _____
1. _____
2. _____
3. _____
Diagnostic test _____
Breath hose heated _____
Printer check _____

REMARKS:

Intox. Ser. # ⁰⁰68-011779 reported o/o/s due to out of tolerance. Intox. Ser. #68-013546 replaced Intox. ⁰¹¹⁷⁷⁹68-013546. Sim. Ser. # DR4147 replaced Sim. Ser. # DR4153. Hwy Dist. notified / PO Seniak present.

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 10020/2012 (N.Y. City S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779
SLAVE 75 2240
DATE 06/18/2008
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 19:31 EDT

6/18/08

OUT OF SERVICE

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779
SLAVE 75_2240
DATE 06/18/2008
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 19:48 EDT

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679 - 1990

SN 68-010779
SLAVE 75_2240
DATE 06/18/2008
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 19:51 EDT

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679 - 1990