



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor G. K. Hannig
Location 45 PCT
Date of Inspection 11/14/2009

Intoxilyzer Serial # 68-010779 Simulator Serial # OK-3561 Lot # 08300
1. 095
2. 097
3. 098
Diagnostic test OK
Breath hose heated OK
Printer check OK

Intoxilyzer Serial # _____ Simulator Serial # _____ Lot # _____
1. _____
2. _____
3. _____
Diagnostic test _____
Breath hose heated _____
Printer check _____

REMARKS

Intoxilyzer 68-010779 was placed out of service due to HIV exposure by Highway district. Said instrument was sanitized and continuous purge was run. The instrument was placed back into service Highway District notified.

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
01/14/2009

DIAGNOSTIC TEST 15:22 EST

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC	PASSED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMNOPQRSTUVWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEUYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
01/14/2009

TEST	%BrAC	TIME
AIR BLANK	.000	15:37EST
CAL. CHECK	.095	15:38EST
AIR BLANK	.000	15:38EST
CAL. CHECK	.097	15:38EST
AIR BLANK	.000	15:38EST
CAL. CHECK	.098	15:39EST
AIR BLANK	.000	15:39EST
NO. OF SAMPLES:	03	
MEAN:	.0966	
STD. DEVIATION:	.0015296	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779
SLAVE 75_2240
DATE 01/14/2009
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^&*()

VERSION G1776.12

TIME 15:39 EST

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779

VERSION G1776.12

SLAVE 75 2240

DATE 01/14/2009

TIME 15:39 EST

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[] ; , . / < > ? abcdefghijklmno
pqrstuvwxyz!@#\$%^*()

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990