


The City of New York



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT  
INTOXILYZER 5000EN

Technical Supervisor

G. K. Hanna

Location

45 PCT

Date of Inspection

3/11/09

Intoxilyzer Serial #

68-010780

Simulator Serial #

DR2679

Lot #

09020

1.

.101

2.

.102

3.

.102

Diagnostic test

OK

Breath hose heated

OK

Printer check

OK

Intoxilyzer Serial #

68-010779

Simulator Serial #

DR3561

Lot #

09020

1.

.097

2.

.098

3.

.099

Diagnostic test

OK

Breath hose heated

OK

Printer check

OK

REMARKS:

Intoxilyzer Ser. # 68-010779 was placed o/o/s by Hwy district on 3/7/09 due to "HIV Exposure". Said instrument was disinfected, fused and returned to service on 3/11/09, notified Hwy district.

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. City. S. Ct.) WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010779  
03/11/2009

DIAGNOSTIC TEST 12:13 EST

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC PASSED

PRINTER CHECK  
ABCDEFGHIJKLMNQPQRSTUVWXYZ  
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of William D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010779  
03/11/2009

TEST	%BrAC	TIME
AIR BLANK	.000	12:14EST
CAL. CHECK	.097	12:14EST
AIR BLANK	.000	12:14EST
CAL. CHECK	.098	12:15EST
AIR BLANK	.000	12:15EST
CAL. CHECK	.099	12:15EST
AIR BLANK	.000	12:16EST
NO. OF SAMPLES:	03	
MEAN:	.0980	
STD. DEVIATION:	.0010000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779  
SLAVE 75 2240  
DATE 03/11/2009  
1234567890ABCDEFGHIJKLMN OPQRSTUVWXYZ-=[ ]; ,./<>?abcdefghijklmnop  
pqrstuvwxyz!@#\$%^\*()

VERSION G1776.12

TIME 12:16 EST

045 PCT

\*\* CURRENT INSTRUMENT SETUP \*\*

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

IPtoX # 68-010 779

O/C/S as of

3/7/09 @ 0530 hrs Due to

HIV Exposure

New York Dist.

PO Bent Notified

- PO Skape

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. N.Y.P.D. et al., Case No. 100220/2012 (N.Y. Civ. S. Ct.). WWW.NEWMYORKLEGALDEFENSE.COM - (212) 679-1990