



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor AA

Location 45 PCT

Date of Inspection 8/17/09

Intoxilyzer Serial # 68-010780 Simulator Serial # DR-2682 Lot # 09150

- 1. .100
- 2. .101
- 3. .100

Diagnostic test OK
 Breath hose heated OK
 Printer check OK

Intoxilyzer Serial # 68-010779 Simulator Serial # DR3561 Lot # 09150

- 1. .095
- 2. .096
- 3. .095

Diagnostic test OK
 Breath hose heated OK
 Printer check OK

REMARKS:

Simulator DR2682 replaced simulator DR 2679 due to agitator paddle not working. Simulator solution was replaced with new solution, lot # 09150.

Intoxilyzer 68-010779 was placed out of service due to diagnostic failure. Upon inspection instrument found to be in good working order. Simulator solution in simulator DR3561 was replaced with new solution, lot # 09150.

Both instruments placed back into service Highway District notified.

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Adrian Arav, Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York City Police Laboratory which has been delegated to my possession, custody and control, by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection Report of Intoxilyzer, Serial Number, 68-010779 conducted on August 17, 2009 by Adrian Arav Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date August 17, 2009

(s) 
Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McGinness, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S.Ct.)
WWW.NYOKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
08/17/2009

DIAGNOSTIC TEST 17:38 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC	PASSED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMNORSTUVWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEUYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
08/17/2009

TEST	%BrAC	TIME
AIR BLANK	.000	17:35EDT
CAL. CHECK	.095	17:36EDT
AIR BLANK	.000	17:36EDT
CAL. CHECK	.096	17:36EDT
AIR BLANK	.000	17:36EDT
CAL. CHECK	.095	17:37EDT
AIR BLANK	.000	17:37EDT
NO. OF SAMPLES:	03	
MEAN:	.0953	
STD. DEVIATION:	.0005742	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010779
SLAVE 75_2240
DATE 08/17/2009
1234567890ABCDEFGHIJKLMN0PQRSTUVWXYZ--[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^*()

VERSION G1776.12

TIME 17:39 EDT

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
08/15/2009

DIAGNOSTIC TEST 23:45 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	FAILED

DIAGNOSTIC	FAILED
------------	--------

PRINTER CHECK
ABCDEFGHIJKLMN OPQRSTU VWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORILEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010779
08/17/2009

TEST	%BrAC	TIME
AIR BLANK	.000	18:11EDT
INTERNAL 1	.100	18:11EDT
INTERNAL 2	.190	18:11EDT
INTERNAL 3	.289	18:11EDT
AIR BLANK	.000	18:12EDT

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990