



POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Adrian Arav, Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518 subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York City Police Laboratory which has been delegated to my possession, custody and control, by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection Report of Intoxilyzer, Serial Number, 68-010779 conducted on August 20, 2009 by Adrian Arav Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date August 20, 2009

(s) [Signature]

Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 10022/2012 (N.Y. Cty. S.Ct.) WWW.NEWTORKLEGALDEFENSE.COM - (212) 679-1990

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor AA

Location 45 PCT

Date of Inspection 8/20/09

Intoxilyzer Serial # 68-010779 Simulator Serial # DR-3561 Lot # 09150

- 1. /
- 2. /
- 3. /

Diagnostic test Failed
Breath hose heated OK
Printer check OK

Intoxilyzer Serial # / Simulator Serial # / Lot # /

- 1. /
- 2. /
- 3. /

Diagnostic test /
Breath hose heated /
Printer check /

REMARKS:

Intoxilyzer 68-010779 was placed out of service for failing a diagnostic test and was replaced by intoxilyzer 68-011087.

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990