



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor P. Zippo
Location 45 Pct.
Date of Inspection 1/29/08 *(Signature)*

Intoxilyzer Serial # 68-010782 Simulator Serial # DR448 Lot # 07250
1. .096
2. .095
3. .096
Diagnostic test OK
Breath hose heated OK
Printer check OK

Intoxilyzer Serial # 68-010780 Simulator Serial # DR3561 Lot # 07250
1. .099
2. .098
3. .099
Diagnostic test OK
Breath hose heated OK
Printer check OK

REMARKS:

Intox. Ser # 68-010780 reported 0/0/5 by Hwy Dist. as of 1/29/08. on 1/29/08 for unknown reason. Found instrument on, simulator turned off. Powered off instrument, turned back on. Turned simulator on. Proceeded with diagnostics, Cal. OK. Instrument working properly. Instrument returned back to service. Hwy District notified *(Signature)*. PO Serial.

Obtained by Perlmutter, P.C. v. NYPD, et al.
Case No. 100220/2014 (N.Y. Crim. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD et al. NYPD
Case No. 100220/2012 (N.Y. City. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 979-1100

Q

U T

S

E R

80/22/108

Intr. Ser. # 68010780

Simulster found turned off; ibralox

" SHARPE IS MINT "

45 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
01/29/2008

DIAGNOSTIC TEST 21:24 EST

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240

RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMNQPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of William D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

45 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
01/29/2008

TEST	%BrAC	TIME
AIR BLANK	.000	21:25EST
CAL. CHECK	.099	21:25EST
AIR BLANK	.000	21:26EST
CAL. CHECK	.098	21:26EST
AIR BLANK	.000	21:26EST
CAL. CHECK	.099	21:27EST
AIR BLANK	.000	21:27EST
NO. OF SAMPLES:	03	
MEAN:	.0986	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780

VERSION G1776.12

SLAVE 75_2240

DATE 01/29/2008

TIME 21:28 EST

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^*()

45 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990