



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT  
INTOXILYZER 5000EN

Technical Supervisor P. Zippo  
Location 45 Pct.  
Date of Inspection 6/27/08

Intoxilyzer Serial # 68-010779 Simulator Serial # DL3561 Lot # 08070

- 1. .101
- 2. .103
- 3. .103

Diagnostic test OK  
Breath hose heated OK  
Printer check OK

Intoxilyzer Serial # 68-01196 Simulator Serial # DR2679 Lot # 08070

- 1. .101
- 2. .103
- 3. .102

Diagnostic test OK  
Breath hose heated OK  
Printer check OK

REMARKS:

Intox. Ser. # 68-010780 o/o/s due to Int. Std. / Diagnostic fail. Replaced with Intox. Ser. # 68-010779. Changed 1070 Ref EtOH Sol'n. in Sim. Ser. # DR3561 to Lot # 08070, changed 1070 Ref. EtOH Sol'n. in Sim. Ser. # DR2679 to Lot # 08

Intox. Ser. # 68-01196 o/o/s due to out of tolerance. Reset target value from .080 to .100. Performed Dian. / Cal. ck. - Field Inspection. Ret'd back to service.

Signed - PO Phillips (Signature)

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 1002207212 (NY Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

45 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
06/22/2008

DIAGNOSTIC TEST 05:17 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	FAILED

DIAGNOSTIC	FAILED
------------	--------

PRINTER CHECK  
ABCDEFGHIJKLMNPOQRSTUVWXYZ  
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

45 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
06/22/2008

DIAGNOSTIC TEST 05:18 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	FAILED

DIAGNOSTIC	FAILED
------------	--------

PRINTER CHECK  
ABCDEFGHIJKLMNPOQRSTUVWXYZ  
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

45 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
06/22/2008

DIAGNOSTIC TEST 05:23 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	FAILED

DIAGNOSTIC	FAILED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMNPOQRSTUVWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

45 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
06/22/2008

DIAGNOSTIC TEST 05:07 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	FAILED

DIAGNOSTIC	FAILED
------------	--------

PRINTER CHECK  
ABCDEFGHIJKLMNORSTUVWXYZ  
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

F

O

LO

Index Ser #  
68-010780  
C

6/22/08

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C., NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990