



POLICE DEPARTMENT

FIELD UNIT INSPECTION REPORT
INTOXILYZER 5000EN

Technical Supervisor G.K. Hanna

Location 2 45 PCT

Date of Inspection 2/17/09

Intoxilyzer Serial # 68-010779 Simulator Serial # DR3561 Lot # 08300

- 1. .096
- 2. .097 (E)
- 3. .097

Diagnostic test OK
 Breath hose heated OK
 Printer check OK

Intoxilyzer Serial # 68-010780 Simulator Serial # DR2679 Lot # 08300

- 1. .100
- 2. .099 (E)
- 3. .100

Diagnostic test OK
 Breath hose heated OK
 Printer check OK

REMARKS:

Intoxilyzer serial # 68-010780 was placed out of service for no specified reason, a field breath test was performed with good readings and placed back in service. PO Randazzo from Hwy District was notified @ 1614 hrs

Obtained by Perlmutter & McQuinn, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al., WWW.NEYORLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
02/17/2009

DIAGNOSTIC TEST 15:54 EST

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240

RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMN OPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NYOSKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
02/17/2009

TEST	%BrAC	TIME
AIR BLANK	.000	15:54EST
CAL. CHECK	.100	15:55EST
AIR BLANK	.000	15:55EST
CAL. CHECK	.099	15:55EST
AIR BLANK	.000	15:55EST
CAL. CHECK	.100	15:56EST
AIR BLANK	.000	15:56EST
NO. OF SAMPLES:	03	
MEAN:	.0996	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75_2240
DATE 02/17/2009

VERSION G1776.12

TIME 16:00 EST

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ--[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^*()

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
02/17/2009

TEST	%BrAC	TIME
AIR BLANK	.000	15:57EST
CAL. CHECK	.100	15:57EST
AIR BLANK	.000	15:57EST
CAL. CHECK	.100	15:58EST
AIR BLANK	.000	15:58EST
CAL. CHECK	.100	15:58EST
AIR BLANK	.000	15:58EST
CAL. CHECK	.100	15:59EST
AIR BLANK	.000	15:59EST
CAL. CHECK	.100	15:59EST
AIR BLANK	.000	16:00EST
NO. OF SAMPLES:	05	
MEAN:	.1000	
STD. DEVIATION:	.0000000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDISPUTE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
02/17/2009

TEST	%BrAC	TIME
AIR BLANK	.000	15:57EST
CAL. CHECK	.100	15:57EST
AIR BLANK	.000	15:57EST
CAL. CHECK	.100	15:58EST
AIR BLANK	.000	15:58EST
CAL. CHECK	.100	15:58EST
AIR BLANK	.000	15:58EST
CAL. CHECK	.100	15:59EST
AIR BLANK	.000	15:59EST
CAL. CHECK	.100	15:59EST
AIR BLANK	.000	16:00EST
NO. OF SAMPLES:	05	
MEAN:	.1000	
STD. DEVIATION:	.0000000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam B. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
02/17/2009

I.D.T.U. TEST # 1

DEFENDANT

NAME: Q,Q,Q
ADDRESS: A
DRIVER LIC# 1/NY

AGE 41 SEX M
DATE OF BIRTH 02/06/68
RACE W

OCCURANCE

PLACE: Y
CHECK POINT N
ARR OFF OBS Y

TIME: 12:00 ACCIDENT N
PATROL STOP N NUMBER KILLED
CIVILIAN OBS N NUMBER INJURED

ARREST

TIME OF ARREST 12:30 TOP CHARGE
ARREST OFFICER: A,A

SHIELD 1 ADD CHARGES
TAX NUMBER 1 COMMAND 79

FIELD TEST N

TEST

BREATH TEST Y BLOOD TEST N
SIMULATOR SER# DR2679
TECH NAME : A,S

URINE TEST N ADD TEST N
STD SOL CONTL# 08300
SHIELD 1 COMMAND 79
TAX NUMBER 1

TEST	%BrAC	TIME
AIR BLANK	.000	16:03EST
CAL. CHECK	.101	16:03EST
AIR BLANK	.000	16:04EST
SUBJECT TEST	.000	16:04EST
AIR BLANK	.000	16:04EST

REFUSAL N INVALID SAMPLE N
BREATH INTERFERENT N

INSUFFICIENT SAMPLE N
CAL. CHECK INTERFERENT N

VEHICLE

CAR SEIZED N

WHY NOT OWNER
CAR PLATE

Obtained by Perlmutter & McGrath, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. N.Y.C. Police Dept., et al.,
Case No. 100220/2012 (N.Y. City Ct.). WWW.NEWMYORKLEGALDEFENSE.COM - (212) 679-1990

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Ct. App. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

087010 - 89 #

82H 02 to

60/5/09

5/0/0