

FIELD UNIT INSPECTION REPORT  
INTOXILYZER 5000EN

Technical Supervisor G.K. Hanne

Location 45 PCT

Date of Inspection 1/22/10

Intoxilyzer Serial # 68-010780 Simulator Serial # DR4427 Lot # 09260

1. .098
2. .099
3. .099

Diagnostic test OK

Breath hose heated OK

Printer check OK

Intoxilyzer Serial # 68-011087 Simulator Serial # DR3549 Lot # 09260

1. .098
2. .100
3. .098

Diagnostic test OK

Breath hose heated OK

Printer check OK

REMARKS:

Both Intoxilyzers Ser. # 68-010780 and 68-011087 were reported o/p/s by Hwy district on 1/22/10 due to "Out of Tolerance". Said instruments were inspected and found to be in proper working condition, Returned back in service notified Hwy district at appx 1545 on 1/22/10.

Obtained by Perlmutter & McGinness, P.C. in New York City, N.Y. City, S. Ct.). WWW.NYORKLEGALDEFENSE.COM - (212) 679-1990  
Perlmutter, P.C. v. NYPD, et al.,



POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY  
OF FIELD UNIT INSPECTION REPORT  
(CPLR RULE 4518)

I, Gamil Hanna Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of The original record of the New York City Police Laboratory which has been delegated to My possession, custody and control by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection Report of Intoxilyzer, Serial Number, 68-010780 conducted on January 22, 2010 by Gamil Hanna Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date January 25, 2010

(s)

Technical Supervisor  
Intoxicated Drivers Testing Unit  
New York City Police Laboratory

Obtained by Perlmutter & McGinness, P.C. in law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. S.C.). WWW.NEWCYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
01/22/2010

DIAGNOSTIC TEST 15:30 EST

PROM CHECK PASSED  
Z80 VER - G1776.12  
SLAVE 75\_2240  
RAM CHECK PASSED  
TEMP CHECK PASSED  
PROCESSOR CHECK  
MOTOR CHECK PASSED  
EEPROM CHECK PASSED  
SERIAL NO. MATCH PASSED  
RANGE/STABILITY PASSED  
AUTO CAL STATUS PASSED

RTC CHECK PASSED  
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK  
ABCDEFGHIJKLMNORSTUVWXYZ  
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
01/22/2010

TEST	%BrAC	TIME
AIR BLANK	.000	15:31EST
CAL. CHECK	.098	15:31EST
AIR BLANK	.000	15:31EST
CAL. CHECK	.099	15:32EST
AIR BLANK	.000	15:32EST
CAL. CHECK	.099	15:32EST
AIR BLANK	.000	15:32EST
NO. OF SAMPLES:	03	
MEAN:	.0986	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780  
SLAVE 75\_2240  
DATE 01/22/2010  
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[ ]; ,./<>?abcdefghijklmnop  
pqrstuvwxyz!@#\$%^\*()

VERSION G1776.12

TIME 15:33 EST

045 PCT

\*\* CURRENT INSTRUMENT SETUP \*\*

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780  
SLAVE 75\_2240  
DATE 01/22/2010  
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[ ]; ,./<>?abcdefghijklmnop  
pqrstuvwxyz!@#\$%^\*()

VERSION G1776.12

TIME 15:33 EST

045 PCT

\*\* CURRENT INSTRUMENT SETUP \*\*

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990