



POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT

Instrument INTOXILIZER 5000EN Date of Inspection 7/6/10

Location 45 PCT

Intoxilizer Serial # 68-010780 Simulator Serial # DR4427 Lot # 00000

1. .102
2. .102
3. .101

Diagnostic test OK
Breath tube heated OK
Printer check OK

Intoxilizer Serial # _____ Simulator Serial # _____ Lot # _____

1. _____
2. _____
3. _____

Diagnostic test _____
Breath tube heated _____
Printer check _____

Remarks Intoxilizer 68-010780 was placed out of service by Highway District due to simulator malfunction. Upon inspection said instrument and simulator were found to be in good working order and placed back into service Highway District notified.

PO ARAV POTA
Rank / Name / Signature
Technical Supervisor

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 10020/2012 (N.Y. Cty. S. Ct.). WWW.NYORKLEGALDEFENSE.COM - (212) 679-1990




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Gamil Hanna Technical Supervisor, Intoxicated Drivers Testing Unit,
New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-
3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c,
Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of
The original record of the New York City Police Laboratory which has been delegated to
My possession, custody and control by the Director of the New York City Police
Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer,
Serial number, 68-010780 conducted on July 6, 2010 by Adrian Arav
Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a
photocopy, was made in the regular course of business of the New York City Police
Laboratory, and that it was the regular course of New York City Police Laboratory
business to make such record at the time of each event recorded in it or within a
reasonable time thereafter.

Date July 12, 2010

(s) 

Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McQuinn, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. City S. Ct.) WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
07/06/2010

DIAGNOSTIC TEST 16:44 EDT

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMN OPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
07/06/2010

TEST	%BrAC	TIME
AIR BLANK	.000	16:41EDT
CAL. CHECK	.102	16:41EDT
AIR BLANK	.000	16:41EDT
CAL. CHECK	.102	16:42EDT
AIR BLANK	.000	16:42EDT
CAL. CHECK	.101	16:42EDT
AIR BLANK	.000	16:42EDT
NO. OF SAMPLES:	03	
MEAN:	.1016	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
07/06/2010

I.D.T.U. TEST # T

DEFENDANT

NAME: T,T,G AGE 41 SEX M
ADDRESS: K DATE OF BIRTH 09/09/96
DRIVER LIC# 4/Y RACE W

OCCURANCE

PLACE: T TIME: 12:00 ACCIDENT N
CHECK POINT N PATROL STOP N NUMBER KILLED
ARR OFF OBS Y CIVILIAN OBS N NUMBER INJURED

ARREST

TIME OF ARREST 12:09 TOP CHARGE ADD CHARGES
ARREST OFFICER: W,Q SHIELD 123 COMMAND 12
FIELD TEST N TAX NUMBER 123

TEST

BREATH TEST Y BLOOD TEST N URINE TEST N ADD TEST N
SIMULATOR SER# DR4427 STD SOL CONTL# 10060
TECH NAME : P,P SHIELD 987 COMMAND 79
TAX NUMBER 987

TEST	%BrAC	TIME
AIR BLANK	.000	16:27EDT
CAL. CHECK	.098	16:27EDT
AIR BLANK	.000	16:27EDT
SUBJECT TEST	.000	16:28EDT
AIR BLANK	.000	16:28EDT

REFUSAL N INVALID SAMPLE N INSUFFICIENT SAMPLE N
BREATH INTERFERENT N CAL. CHECK INTERFERENT N

VEHICLE

CAR SEIZED N WHY T
CAR PLATE

Obtained by Perlmutter & McQuinn, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Sup. Ct.). WWW.NEUYORKLEGALDEFENSE.COM (212) 679-1990

SN 68-010780
SLAVE 75_2240
DATE 07/06/2010
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$%^*()

VERSION G1776.12

TIME 16:44 EDT

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990