



POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT

Instrument INTOXILIZER 5000EN Date of Inspection 8/19/10

Location 45 pct.

Intoxilizer Serial # 68-010780 Simulator Serial # DR4381 Lot # 10060

- 1. .102
- 2. .103
- 3. .103

Diagnostic test OK
Breath tube heated OK
Printer check OK

Intoxilizer Serial # 68-011087 Simulator Serial # DR3549 Lot # 10060

- 1. .100
- 2. .101
- 3. .102

Diagnostic test OK
Breath tube heated OK
Printer check OK

Remarks Intoxilizer Ser. #68-010780 was placed P/O/S by Hwy-1 on 8/18/10 due to "HIV Contamination". Lab/IDU was not notified. Said instrument was disinfected, purged and returned to service. Hwy district was notified at appx 1145hrs on 8/19/10

Crim III G.K. Hanna

Rank / Name / Signature
Technical Supervisor

Obtained by Perlmutter & McGinnese (N.Y. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990
P.C. v. NYPD, et al.




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Gamil Hanna Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of The original record of the New York City Police Laboratory which has been delegated to My possession, custody and control by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer, Serial number, 68-010780 conducted on August 19, 2010 by Gamil Hanna Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date August 19, 2010

(s) 

Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. City. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
08/19/2010

DIAGNOSTIC TEST 11:14 EDT

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMNOPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
08/19/2010

TEST	%BrAC	TIME
AIR BLANK	.000	11:15EDT
CAL. CHECK	.102	11:16EDT
AIR BLANK	.000	11:16EDT
CAL. CHECK	.103	11:16EDT
AIR BLANK	.000	11:17EDT
CAL. CHECK	.103	11:17EDT
AIR BLANK	.000	11:17EDT
NO. OF SAMPLES:	03	
MEAN:	.1026	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75 2240
DATE 08/19/2010
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#%&^*()

VERSION G1776.12

TIME 11:19 EDT

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

08/08/2010

@ 1945 #

0/0/5

HIV

subject
=

~~IN~~ TU #

2010 BX 993

O.O

Marshall

H-4-1

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S.Ct.)
WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990