



POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT

Instrument INTOXILIZER 5000EN

Date of Inspection 4/18/11

Location 45 PCT

Intoxilizer Serial # 68-010780

Simulator Serial # DR4381

Lot # 11010

1. .101

2. .100

3. .101

Diagnostic test OK

Breath tube heated OK

Printer check OK

Intoxilizer Serial # 68-011087

Simulator Serial # DR3549

Lot # 11010

1. .103

2. .103


3. .102

Diagnostic test OK

Breath tube heated OK

Printer check OK

Remarks Intoxilizer 68-010780 placed out of service due to acetone methanol testing and only printing Interferent detected and aborting test. Highway District notified.

PO ARAV 

Rank / Name / Signature
Technical Supervisor

Obtained by Permuter & McGinness, P.C. in Law Offices of Adam D. Permuter, P.C. v. NYPD, et al., Case No. 10022012 (N.Y. Ct. S. Ct.). WWW.NEWMYORKLEGALDEFENSE.COM - (212) 679-1990



POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Adrian Arav Technical Supervisor, Intoxicated Drivers Testing Unit,
New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-
3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c,
Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of
The original record of the New York City Police Laboratory which has been delegated to
My possession, custody and control by the Director of the New York City Police
Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer,
Serial number, 68-010780 conducted on April 18, 2011 by Adrian Arav Technical
Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy,
was made in the regular course of business of the New York City Police Laboratory, and
that it was the regular course of New York City Police Laboratory business to make such
record at the time of each event recorded in it or within a reasonable time thereafter.

Date April 18, 2011

(s) 

Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McGinniss, P.C. in *Law Offices of Adair D. Perlmutter, P.C. v. NYPD, et al.*,
Case No. 100220/2012 (N.Y. Cty. S.C.) WWW.NEWCYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/18/2011

DIAGNOSTIC TEST 15:56 EDT

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMNOPQRSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORILEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/18/2011

TEST	%BrAC	TIME
AIR BLANK	.000	15:56EDT
CAL. CHECK	.101	15:57EDT
AIR BLANK	.000	15:57EDT
CAL. CHECK	.100	15:57EDT
AIR BLANK	.000	15:58EDT
CAL. CHECK	.101	15:58EDT
AIR BLANK	.000	15:58EDT
NO. OF SAMPLES:	03	
MEAN:	.1006	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780

VERSION G1776.12

SLAVE 75 2240

DATE 04/18/2011

TIME 16:30 EDT

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop

qrstuvwxyz!@#%&^*()

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990