



**POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT**

Instrument INTOXILIZER 5000EN Date of Inspection 4/21/11

Location 45 PCT

Intoxilizer Serial # 68-011087 Simulator Serial # DR3549 Lot # _____

- 1. .100
- 2. .100
- 3. .100

Diagnostic test OK
Breath tube heated OK
Printer check OK

Intoxilizer Serial # 68-010780 Simulator Serial # DR 4381 Lot # 11010


- 1. .099
- 2. .098
- 3. .099

Diagnostic test OK
Breath tube heated OK
Printer check OK

Remarks Intoxilizer 68-010780 was placed out of service due to acetone/ethanol testing and only printing Interferent Detected and aborting test. New acetone/ethanol solution prepared instrument able to print proper results of ethanol concentration. Instrument placed back in service Highway District 2011.

POARAU POA
Rank / Name / Signature
Technical Supervisor

Obtained by Permuter & McGinness, P.C. in Law Offices of Adam D. Permuter, P.C. v. NYPD, et al., Case No. 10-220-2012 (N.Y. City. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Adrian Arav Technical Supervisor, Intoxicated Drivers Testing Unit,
New York City Police Laboratory, 150-14 Jamaica Avenue, Queens, New York 11432-
3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c,
Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of
The original record of the New York City Police Laboratory which has been delegated to
My possession, custody and control by the Director of the New York City Police
Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer,
Serial number, 68-010780 conducted on April 21, 2011 by Adrian Arav Technical
Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy,
was made in the regular course of business of the New York City Police Laboratory, and
that it was the regular course of New York City Police Laboratory business to make such
record at the time of each event recorded in it or within a reasonable time thereafter.

Date April 21, 2011

(s) 

Technical Supervisor
Intoxicated Drivers Testing Unit
New York City Police Laboratory

Obtained by Perlmutter & McGinniss, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. D.C.) WWW.PERMUTTER.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/21/2011

DIAGNOSTIC TEST 10:38 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC	PASSED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMNORSTUVWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/21/2011

TEST	%BrAC	TIME
AIR BLANK	.000	10:39EDT
CAL. CHECK	.099	10:39EDT
AIR BLANK	.000	10:39EDT
CAL. CHECK	.098	10:40EDT
AIR BLANK	.000	10:40EDT
CAL. CHECK	.099	10:40EDT
AIR BLANK	.000	10:41EDT
NO. OF SAMPLES:	03	
MEAN:	.0986	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780

VERSION G1776.12

SLAVE 75 2240

DATE 04/21/2011

TIME 10:46 EDT

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop

pqrstuvwxyz!@#%&^*()

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990