



POLICE LABORATORY  
FIELD UNIT  
INSPECTION REPORT

Instrument INTOXILYZER 5000EN Date of Inspection 1/9/12

Location 45 PCT

Intoxilyzer Serial # 68-010780 Simulator Serial # DR 4381 Lot # 11190

- 1. .100
- 2. .100
- 3. .100

Diagnostic test ok  
Breath tube heated ok  
Printer check ok

Intoxilyzer Serial # 68-011477 Simulator Serial # DR 3549 Lot # 11190

- 1. .100
- 2. .101
- 3. .101

Diagnostic test ok  
Breath tube heated ok  
Printer check ok

Remarks Both Intoxilyzers 68-010780 and 68-011477 were placed out of service by Highway District due to HIV exposure. Both instruments were cleaned and passed and placed back into service, Highway District notified. Simulator solution on both simulators has been replaced with new solution, new solution lot #: 11190.

PO ARAU POA  
Rank / Name / Signature  
Technical Supervisor

Obtained by Permutter & McGuinness, P.C. in Law Offices of Adam D. Permutter, P.C. v. NYPD, et al., Case No. 1002012012 (N.Y. Cty. S. Ct.). WWW.NYORKLEGALDEFENSE.COM - (212) 679-1990




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY  
OF FIELD UNIT INSPECTION REPORT  
(CPLR RULE 4518)

I, Richard Distler, Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Jamaica, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York City Police Laboratory, which has been delegated to My possession, custody and control by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer 5000EN serial number 68-010780, conducted on January 9, 2012 by Adrian Arav, Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date January 10, 2012

(s) 

Technical Supervisor  
Intoxicated Driver Testing Unit  
New York City Police Laboratory

Obtained by Permutter & McGuinness, P.C. in Law Offices of Alan D. Permutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NYKLE.GALLERY.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
01/09/2012

DIAGNOSTIC TEST 15:10 EST

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC	PASSED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMNOPQRSTUVWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT  
INTOXILYZER - ALCOHOL ANALYZER  
MODEL 5000EN SN 68-010780  
01/09/2012

TEST	%BrAC	TIME
AIR BLANK	.000	15:11EST
CAL. CHECK	.100	15:11EST
AIR BLANK	.000	15:11EST
CAL. CHECK	.100	15:11EST
AIR BLANK	.000	15:12EST
CAL. CHECK	.100	15:12EST
AIR BLANK	.000	15:12EST
NO. OF SAMPLES:	03	
MEAN:	.1000	
STD. DEVIATION:	.0000000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780

DATE 01/09/2012

1234567890ABCDEFGHIJKLMNPOQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop

pqrstuvwxyz!@#\$%^\*()

VERSION G1776.12

TIME 15:42 EST

045 PCT

\*\* CURRENT INSTRUMENT SETUP \*\*

SETUP WORD = 0x0029

STD TEST = CUST 1

3 DIGITS ON? Y/N = TRUE

PRELIM RES? Y/N = FALSE

DATA ENTRY? Y/N = TRUE

PRINT INHIB? Y/N = FALSE

SAMPLE CAPT? Y/N = FALSE

INT STDS? Y/N = FALSE

PRINT VOLUME Y/N = FALSE

SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,  
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990