



POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT

Instrument INTOXILYZER 5000EN

Date of Inspection 3/3/12

Location 45 PCT

Intoxilyzer Serial # 68-010780

Simulator Serial # DR 4381

Lot # 11280

1. .103
2. .103
3. .103

Diagnostic test ok

Breath tube heated ok

Printer check ok

Intoxilyzer Serial # 68-011477

Simulator Serial # DR 3549

Lot # 11240

1. .101
2. .101
3. .102

Diagnostic test ok

Breath tube heated ok

Printer check ok

Remarks Intoxilyzer 68-010780 was placed out of service due to
reference check out of tolerance. Upon inspection serial instrument was
found to be in good working order and placed back into service
Hudson district notified.

PO ARAV POA

Rank / Name / Signature
Technical Supervisor




POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Richard Distler, Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Jamaica, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York City Police Laboratory, which has been delegated to My possession, custody and control by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer 5000EN serial number 68-010780, conducted on March 3, 2012 by Adrian Arav, Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date March 5, 2012

(s) 

Technical Supervisor
Intoxicated Driver Testing Unit
New York City Police Laboratory

Obtained by Permutter & McGinness, P.C. in Law Office of Adam D. Permutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
03/03/2012

DIAGNOSTIC TEST 10:12 EST

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240

RAM CHECK PASSED

TEMP CHECK PASSED

PROCESSOR CHECK

MOTOR CHECK PASSED

EEPROM CHECK PASSED

SERIAL NO. MATCH PASSED

RANGE/STABILITY PASSED

AUTO CAL STATUS PASSED

RTC CHECK PASSED

INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK

ABCDEFGHIJKLMNOPQRSTUVWXYZ

0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
03/03/2012

TEST	%BrAC	TIME
AIR BLANK	.000	10:25EST
CAL. CHECK	.103	10:26EST
AIR BLANK	.000	10:26EST
CAL. CHECK	.103	10:26EST
AIR BLANK	.000	10:26EST
CAL. CHECK	.103	10:27EST
AIR BLANK	.000	10:27EST
NO. OF SAMPLES:	03	
MEAN:	.1030	
STD. DEVIATION:	.0000000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75_2240
DATE 03/03/2012
1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#%&^*()

VERSION G1776.12

TIME 10:34 EST

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
03/03/2012

I.D.T.U. TEST # TEST

DEFENDANT

NAME: T,T,T AGE 44 SEX M
ADDRESS: N DATE OF BIRTH 03/04/68
DRIVER LIC# 1/NY RACE W

OCCURANCE

PLACE: T TIME: 12:00 ACCIDENT N
CHECK POINT N PATROL STOP Y NUMBER KILLED
ARR OFF OBS Y CIVILIAN OBS N NUMBER INJURED

ARREST

TIME OF ARREST 12:09 TOP CHARGE ADD CHARGES
ARREST OFFICER: R,R SHIELD 123 COMMAND 79
TAX NUMBER 123
FIELD TEST N

TEST

BREATH TEST Y BLOOD TEST N URINE TEST N ADD TEST N
SIMULATOR SER# DR4381 STD SOL CONTL# 11240
TECH NAME : R,R SHIELD 123 COMMAND 79
TAX NUMBER 123

TEST	%BrAC	TIME
AIR BLANK	.000	10:31EST
CAL. CHECK	.104	10:31EST
AIR BLANK	.000	10:32EST
SUBJECT TEST	.000	10:32EST
AIR BLANK	.000	10:32EST

REFUSAL N INVALID SAMPLE N INSUFFICIENT SAMPLE N
BREATH INTERFERENT N CAL. CHECK INTERFERENT N

VEHICLE

CAR SEIZED N WHY T
CAR PLATE

Obtained by Perlmutter & McQuinn, P.C. in Law Offices of Adam D. Perlmutter, P.C. NY, NY, et al.,
Case No. 100220/2012 (Ch. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75 2240
DATE 03/03/2012
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 04:51 EST

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990