

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/03/2012

DIAGNOSTIC TEST 14:20 EDT

PROM CHECK	PASSED
Z80 VER - G1776.12	
SLAVE 75_2240	
RAM CHECK	PASSED
TEMP CHECK	PASSED
PROCESSOR CHECK	
MOTOR CHECK	PASSED
EEPROM CHECK	PASSED
SERIAL NO. MATCH	PASSED
RANGE/STABILITY	PASSED
AUTO CAL STATUS	PASSED

RTC CHECK	PASSED
INTERNAL STD	PASSED

DIAGNOSTIC	PASSED
------------	--------

PRINTER CHECK	
ABCDEFGHIJKLMN OPQRSTU VWXYZ	
0123456789	

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORILEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75 2240
DATE 04/03/2012
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 14:52 EDT

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.;
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/03/2012

TEST	%BrAC	TIME
AIR BLANK	.000	14:23EDT
CAL. CHECK	.000	14:23EDT
AIR BLANK	.000	14:23EDT
CAL. CHECK	.000	14:24EDT
AIR BLANK	.000	14:24EDT
CAL. CHECK	.000	14:24EDT
AIR BLANK	.000	14:25EDT
NO. OF SAMPLES:	03	
MEAN:	.0000	
STD. DEVIATION:	.0000000	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75 2240
DATE 04/03/2012
INVALID TEST
*REFERENCE CHECK OUT OF TOLERANCE

VERSION G1776.12

TIME 14:52 EDT

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990