



POLICE LABORATORY
FIELD UNIT
INSPECTION REPORT

Instrument INTOXILYZER 5000EN Date of Inspection 4/4/12
Location 45 PCT

Intoxilyzer Serial # 68-010780 Simulator Serial # DR4381 Lot # 12020

- 1. .100
- 2. .101
- 3. .101
- Diagnostic test ok
- Breath tube heated ok
- Printer check ok

Intoxilyzer Serial # 68-011477 Simulator Serial # DR3549 Lot # 12020

- 1. .097
- 2. .098
- 3. .098
- Diagnostic test ok
- Breath tube heated ok
- Printer check ok

Remarks Intoxilyzer 68-010780 was placed out of service due to out of tolerance. Simulator solution in both simulators was replaced with new solution, new solution lot # 12020. Intoxilyzer 68-010780 working properly with excellent calibration readings and was placed back into service Highway District notified.

PO Arav POA
Rank / Name / Signature
Technical Supervisor

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al., Case No. 10020/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990



POLICE DEPARTMENT

CERTIFICATE OF PHOTOSTATIC COPY
OF FIELD UNIT INSPECTION REPORT
(CPLR RULE 4518)

I, Richard Distler, Technical Supervisor, Intoxicated Drivers Testing Unit, New York City Police Laboratory, 150-14 Jamaica Avenue, Jamaica, New York 11432-3725, do hereby certify and authenticate as provided by Rule 4518, subdivision c, Civil Practice Law and Rules, that the copy annexed hereto is an exact photocopy of the original record of the New York City Police Laboratory, which has been delegated to My possession, custody and control by the Director of the New York City Police Laboratory. I further certify that the original Field Unit Inspection report of Intoxilyzer 5000EN serial number 68-010780, conducted on April 4, 2012 by Adrian Arav, Technical Supervisor, Intoxicated Drivers Testing Unit, of which the copy annexed is a photocopy, was made in the regular course of business of the New York City Police Laboratory, and that it was the regular course of New York City Police Laboratory business to make such record at the time of each event recorded in it or within a reasonable time thereafter.

Date April 5, 2012

(s) [Signature]

Technical Supervisor
Intoxicated Driver Testing Unit
New York City Police Laboratory

Obtained by Permutter & McGuinness, P.C. in an Office of Adam B. Permutter, P.C. v. NYPD, et al., Case No. 100220/2012 (N.Y. Cty. P.C.). WWW.NEYDRILLGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/04/2012

DIAGNOSTIC TEST 14:41 EDT

PROM CHECK PASSED
Z80 VER - G1776.12
SLAVE 75_2240
RAM CHECK PASSED
TEMP CHECK PASSED
PROCESSOR CHECK
MOTOR CHECK PASSED
EEPROM CHECK PASSED
SERIAL NO. MATCH PASSED
RANGE/STABILITY PASSED
AUTO CAL STATUS PASSED

RTC CHECK PASSED
INTERNAL STD PASSED

DIAGNOSTIC PASSED

PRINTER CHECK
ABCDEFGHIJKLMNORSTUVWXYZ
0123456789

Obtained by Perlmutter & McGuinness, P.C. in Law Office of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990

045 PCT
INTOXILYZER - ALCOHOL ANALYZER
MODEL 5000EN SN 68-010780
04/04/2012

TEST	%BrAC	TIME
AIR BLANK	.000	14:56EDT
CAL. CHECK	.100	14:56EDT
AIR BLANK	.000	14:56EDT
CAL. CHECK	.101	14:56EDT
AIR BLANK	.000	14:57EDT
CAL. CHECK	.101	14:57EDT
AIR BLANK	.000	14:57EDT
NO. OF SAMPLES:	03	
MEAN:	.1006	
STD. DEVIATION:	.0005828	

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEYORKLEGALDEFENSE.COM - (212) 679-1990

SN 68-010780
SLAVE 75 2240
DATE 04/04/2012

VERSION G1776.12

TIME 15:03 EDT

1234567890ABCDEFGHIJKLMNOPQRSTUVWXYZ-=[]; ,./<>?abcdefghijklmnop
pqrstuvwxyz!@#\$\$%^*()

045 PCT

** CURRENT INSTRUMENT SETUP **

SETUP WORD = 0x0029
STD TEST = CUST 1
3 DIGITS ON? Y/N = TRUE
PRELIM RES? Y/N = FALSE
DATA ENTRY? Y/N = TRUE
PRINT INHIB? Y/N = FALSE
SAMPLE CAPT? Y/N = FALSE
INT STDS? Y/N = FALSE
PRINT VOLUME Y/N = FALSE
SETUP WORD 2 = 0x0000

STANDBY TIMEOUT = 030

CALIBRATION TYPE = WET BATH

Obtained by Perlmutter & McGuinness, P.C. in Law Offices of Adam D. Perlmutter, P.C. v. NYPD, et al.,
Case No. 100220/2012 (N.Y. Cty. S. Ct.). WWW.NEWYORKLEGALDEFENSE.COM - (212) 679-1990